

## Summary Calibration Report

Ace Pipettes, LLC.  
14230 Prospect Point Drive  
Cypress, TX, 77429  
832-295-0882

Date: 04/29/2020  
Work Order: KB0429201  
Company/Inst: WORLDWIDE CLINICAL TRIALS  
Dept:  
Name: Jacqueline King  
PO:

	Serial#	ID1	ID2	Volume "UL"	Make	Model	Cal. Date	Due Date	As Found	As Left	Type
1	HH45467	D3494	CLINICAL	5000	Thermo Fisher	Finpipette 2	04/29/2020	07/31/2020	Pass	Pass	Single
2	806162	D3433	CLINICAL	1000	Vista Lab	MLA	04/29/2020	07/31/2020	Pass	Pass	Single
3	806117	D3435	CLINICAL	1000	Vista Lab	MLA	04/29/2020	07/31/2020	Pass	Pass	Single
4	502419	D3436	CLINICAL	5000	Vista Lab	MLA	04/29/2020	07/31/2020	Pass	Pass	Single
5	502360	D3437	CLINICAL	10000	Vista Lab	MLA	04/29/2020	07/31/2020	Pass	Pass	Single
6	804630	D3431	CLINICAL	500	Vista Lab	MLA	04/29/2020	07/31/2020	Pass	Pass	Single
7	806505	D3460	CLINICAL	5	Vista Lab	MLA	04/29/2020	07/31/2020	Pass	Pass	Single
8	805475	D3432	CLINICAL	50	Vista Lab	MLA	04/29/2020	07/31/2020	Pass	Pass	Single
9	805498	D3434	CLINICAL	50	Vista Lab	MLA	04/29/2020	07/31/2020	Pass	Pass	Single
10	553664	D3404	CLINICAL	50	Thermo Fisher	Finpipette	04/29/2020	07/31/2020	Pass	Pass	Single
11	J27230B	D2801	CLINICAL	5000	Eppendorf	Research Plus	04/29/2020	07/31/2020	Pass	Pass	Single
12	541440553	D2821	CLINICAL	100	VWR	VE	04/29/2020	07/31/2020	Pass	Pass	Single
13	Q56774H	D3479	CLINICAL	1000	Eppendorf	Research Plus	04/29/2020	07/31/2020	Pass	Pass	Single
14	4602806	D2820	CLINICAL	10	Eppendorf	Research	04/29/2020	07/31/2020	Pass	Pass	Single
15	J11662E	D3072	PHARMACY	1000	Eppendorf	Research Plus	04/29/2020	07/31/2020	Pass	Pass	Single
16	J27031D	D2974	PK LAB	1000	Eppendorf	Research Plus	04/29/2020	07/31/2020	Pass	Pass	Single
17	291029A	D3491	PK LAB	1000	Eppendorf	Research Plus	04/29/2020	07/31/2020	Pass	Pass	Single
18	J33381D	D3000	PK LAB	1000	Eppendorf	Research Plus	04/29/2020	07/31/2020	Pass	Pass	Single
19	J33059D	D3226	PK LAB	1000	Eppendorf	Research Plus	04/29/2020	07/31/2020	Pass	Pass	Single
20	J33214D	D3004	PK LAB	1000	Eppendorf	Research Plus	04/29/2020	07/31/2020	Pass	Pass	Single
21	J27038D	D2975	PK LAB	1000	Eppendorf	Research Plus	04/29/2020	07/31/2020	Pass	Pass	Single
22	3074013	D2811	TRAINING	1000	Eppendorf	Research	04/30/2020	07/31/2020	Pass	Pass	Single
23	J33050D	D3005	TRAINING	1000	Eppendorf	Research Plus	04/30/2020	07/31/2020	Pass	Pass	Single
24	3958943	D2871	TRAINING	200	Eppendorf	Research	04/30/2020	07/31/2020	Pass	Pass	Single
25	114336A-100ul	D3076	TRAINING 2-1000	1000	Eppendorf	Rptr Plus	04/30/2020	07/31/2020	Pass	Pass	Single
26	114336A-1ml	D3076	TRAINING	1000	Eppendorf	Rptr Plus	04/30/2020	07/31/2020	Pass	Pass	Single
27	114336A-5ml	D3076	TRAINING	1000	Eppendorf	Rptr Plus	04/30/2020	07/31/2020	Pass	Pass	Single
28	2671387	D2825	PHARMACY	10000	Eppendorf	Research	04/30/2020	07/31/2020	Pass	Pass	Single
29	3319857	D2824	PHARMACY	200	Eppendorf	Research	04/30/2020	07/31/2020	Pass	Pass	Single
30	4609591	D2810	PHARMACY	1000	Eppendorf	Research	04/30/2020	07/31/2020	Pass	Pass	Single
31	4945856	D2823	PHARMACY	20	Eppendorf	Research	04/30/2020	07/31/2020	Pass	Pass	Single
32	2252936-100ul	D3075	PHARMACY	1000	Eppendorf	Rptr Plus	04/30/2020	07/31/2020	Pass	Pass	Single
33	2252936-1ml	D3075	PHARMACY	1000	Eppendorf	Rptr Plus	04/30/2020	07/31/2020	Pass	Pass	Single
34	2252936-5ml	D3075	PHARMACY	1000	Eppendorf	Rptr Plus	04/30/2020	07/31/2020	Pass	Pass	Single
35	L26531E-100ul	D2891	PHARMACY	1000	Eppendorf	Rptr M4	04/30/2020	07/31/2020	Pass	Pass	Single
36	L26531E-1ml	D2891	PHARMACY	1000	Eppendorf	Rptr M4	04/30/2020	07/31/2020	Pass	Pass	Single
37	L26531E-5ml	D2891	PHARMACY	1000	Eppendorf	Rptr M4	04/30/2020	07/31/2020	Pass	Pass	Single
38	N19150G	D3475	PHARMACY	25	Eppendorf	Research Plus	04/30/2020	07/31/2020	Pass	Pass	Single
39	QJ05285	D3498	PHARMACY	1000	Gilson	Microman	04/30/2020	07/31/2020	Pass	Pass	Single
40	QJ05378	D3496	PHARMACY	1000	Gilson	Microman	04/30/2020	07/31/2020	Pass	Pass	Single
41	QJ05280	D3497	PHARMACY	1000	Gilson	Microman	04/30/2020	07/31/2020	Pass	Pass	Single
42	J32974D	D3008	PK LAB	1000	Eppendorf	Research Plus	04/30/2020	07/31/2020	Pass	Pass	Single
43	J33301D	D3002	PK LAB	1000	Eppendorf	Research Plus	04/30/2020	07/31/2020	Pass	Pass	Single
44	J33127D	D3003	PK LAB	1000	Eppendorf	Research Plus	04/30/2020	07/31/2020	Pass	Pass	Single
45	J33213D	D3007	PK LAB	1000	Eppendorf	Research Plus	04/30/2020	07/31/2020	Pass	Pass	Single

Number of pipets:	Single Channel	Multi Ch "8"	Multi Ch "12"
Level 2	72	0	0

Calibrated by: KB  
Balance/s: Recal due:  
B043079758 OCT/31/2020

Procedure/Notes:  
These pipettes were calibrated gravimetrically with water that is no less than 17 MegaOhm/cm3.  
All balances are calibrated against NIST traceable standards and serviced yearly by the manufacturer.  
This document cannot be reproduced partially.

